

AIHA Beryllium Proficiency Analytical Testing (BePAT) Program

(administered under contract with B&W Y12)

Lab ID: 101070 WI Occupational Health Lab

Round 18 2008

January 30, 2009

Sample ID	Results Qualifier	Reported Results (ug)	Round Mean (ug)	Expected (ug)	Recovery (%)	Number of Participants	Standard Deviation	Coefficient of Variation	Rating	Comments
BER1808-166		12	11.176	11.2	107.14%	48	0.567	0.05075	A	
BER1808-167		5.78	5.338	5.427	106.50%	48	0.286	0.05367	A	
BER1808-168		2.12	1.962	2.03775	104.04%	48	0.141	0.07196	A	
BER1808-169		0.61	0.552	0.5678	107.43%	48	0.065	0.1175	A	
Blank1808-170	<	0.0125		0	Not calculated				*	

Rating Key:
 A - Acceptable (Between 75% and 125% of Expected Value)
 L - <75% of Expected Value
 H - >125% of Expected Value
 * - No rating assigned
 NR - Not Reported

