



**WISCONSIN STATE
LABORATORY OF HYGIENE**

Environmental Health Division
2601 Agriculture Drive
P.O. Box 7996
Madison, WI 53707-7996
Phone: (608) 224-6252 · (800) 442-4618
FAX: (608) 224-6213

University of Wisconsin

National Filter Paper Blood Lead PT Program June 2009 Testing Event



July 03, 2009

Dear Participant:

Your individual performance report and statistical summary for the monthly Filter Paper Blood lead proficiency testing event are attached.

As always, please contact us with any questions you might have.


Sincerely,

Noel Stanton
Pb/EPPT Program Supervisor

Pb/EPPT Program, Toxicology Section
State Laboratory of Hygiene
2601 Agriculture Drive
Madison WI 53718

(608) 224-6252
toxpt@mail.slh.wisc.edu





Target and Acceptable Range Information, ug/dL

Group	09FP20	09FP21	09FP22
Refereed Methods			
Target Value	15	1	12
Accepted Result	11-19	0-5	8-16

Method Statistics

Refereed Group	09FP20	09FP21	09FP22
Referee Laboratories (N=12)			
Mean	15.2	1.2	11.6
Standard Deviation	0.9	0.7	0.5
Number of Results *	12	7	12
All Laboratories (N=13)			
Mean*	15.5	1.2	11.7
Standard Deviation*	1.2	0.7	0.7
Participant Results (%Pass)	88	100	100
Ammonium Phosphate Extraction/GFAA (N=4)			
Mean*	16.5	1.2	11.5
Standard Deviation*	0.6	1.0	1.0
Participant Results (%Pass)	100	100	100
Nitric Acid Extraction/GFAA (N=1)			
Mean*	14.0	2.0	12.0
Standard Deviation*	0.0	0.0	0.0
Participant Results (%Pass)	100	100	100
ICP/MS (N=8)			
Mean*	16.5	1.2	11.9
Standard Deviation*	2.7	0.5	0.8
Participant Results (%Pass)	88	100	100

* Statistical outliers and alpha-numeric results (e.g. <X) have been excluded from the calculations

† Lack of statistical group consensus; all results deemed acceptable.

